

CERTIFICATE

CONCERNING THE TYPE TEST OF ENERGY METERS

Manufacturer / Applicant : *JCS Concern Energomera
Russia*

Instrument : *Single phase energy meter, class 1*

type : *CE 07B*
measurement range : *220 V
5(50) A or 10(100) A
50 Hz
Mechanical register*

Period of investigation : *May 2004 till April 2005*


Results of the type test :

The energy meter meets the requirements mentioned in:
- *IEC 62053-21 (2003): "Electricity metering equipment (a.c.) - Static meters for active energy (classes 1 and 2)", requirements for indoor use and class 1, based on a non-recurrent examination.*

The results of the type test are mentioned in KEMA report with reference number 2071533-QUA/CAL 05-1353.

Arnhem, 26 April 2005

KEMA Quality B.V.



R.H. Garszen

Calibration Laboratory